

A sample of this product has been subject to assessment by Nemko, and is hereby confirmed to be in conformity with the applicable European EMC standards referred below. Sufficient evidence of conformity with Article 4 of the EMC Directive (2004/108/EC) has been provided.

Certificate No.: **250817**


Applicant	APEX TECHNOLOGIES LIMITED. Room A 8/F., Excelsior Building ,68-76 Sha Tsui Road, Tsuen Wan N.T. Hong Kong
Manufacturer	APEX TECHNOLOGIES LIMITED. Room A 8/F., Excelsior Building ,68-76 Sha Tsui Road, Tsuen Wan N.T. Hong Kong
Product	Bio-Well Impulse Analyzer
Model	Bio-Well 2013
Rating	5VDC in
Additional information	N/A
Brand Name	N/A
Documentation	Test report No. : 250817
Standards applied	EN 60601-1-2:2007
This certificate is valid until	2014-03-07 + 48 months (provided that all signed certification conditions are complied with, and that possible changes to the product are notified to Nemko for acceptance prior to implementation)

www.nemko.com

This certificate authorizes the manufacturer to affix the below shown Nemko EMC Mark to each conforming product. And provided it is otherwise confirmed that the product also conforms with any other applicable Directives, the manufacturer (or the European authorized representative) may prepare an EC/EEA Declaration of Conformity and affix the below shown CE-marking to each conforming product.

Date of issue 2014-03-07

Revision: None



signature

Daria Liu

